

Sterilization Tray Validation Summary

InstruSafe® Instrument Protection Systems are validated for each sterilization cycle listed below for use with wrap and/or containers as a sterile barrier.

Reference instrusafe.com for the most up to date information.

STERILIZATION CYCLES + STERILE BARRIERS				
Sterilization Cycle	Wrap	InstruSafe Container	Aesculap® Container	Genesis® Container
Steam 4 minute	✓	✓	✓	✓
Ethylene Oxide	✓			✓
STERRAD® 100S Standard	✓		✓	
STERRAD® 100NX® Standard	✓		✓	
STERRAD® 100NX® Express	✓			
STERRAD® 100NX® Flex	✓		✓	
STERRAD® NX® Standard	✓		✓	
STERIS® AMSCO® V-PRO® 1, V-PRO® 1 Plus, V-PRO® maX + V-PRO® maX 2	✓		✓	
Sterizone® VP4	✓		✓	

STERILIZATION BARRIERS + FOOT OPTIONS	
STERILE BARRIER	FOOT OPTION
<p>WRAP</p> 	<p>Rounders Corner Guards -R</p> 
<p>CONTAINER</p> 	<p>Tall Feet -TF</p> 

ISO 13485 : 2016 certified and CMDCAS certified.
Genesis® is a registered trademark of CareFusion Corporation. Aesculap® is a registered trademark of B. Braun Company. STERRAD®, 100S®, 100NX® are registered trademarks of Advanced Sterilization Products (ASP) Division of Ethicon, Inc. STERIS®, AMSCO®, V-PRO® 1, V-PRO® 1 Plus, V-PRO® maX and V-PRO® maX 2 are registered trademarks of STERIS Corporation. InstruSafe® is a registered trademark of Summit Medical. Sterizone® is a registered trademark of Tso3 Inc. ©2022 Summit Medical. All rights reserved.